



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE

United States Patent and Trademark Office

Address: COMMISSIONER FOR PATENTS

P.O. Box 1450

Alexandria, Virginia 22313-1450

www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/510,942	10/08/2004	Bruce K. Jankowski	2780(203-3093)	7109
Covidien 60 Middletown Avenue North Haven, CT 06473			EXAMINER CHEN, VICTORIA W	
			ART UNIT 3739	PAPER NUMBER
			MAIL DATE 05/23/2008	DELIVERY MODE PAPER

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

### Office Action Summary

**Application No.**

10/510,942

**Applicant(s)**

JANKOWSKI, BRUCE K.

**Examiner**

VICTORIA W. CHEN

**Art Unit**

3739

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --  
**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on 08 October 2004.  
2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.  
3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 1-23 is/are pending in the application.  
4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.  
5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.  
6) ☒ Claim(s) 1-23 is/are rejected.  
7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.  
8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.  
10) ☒ The drawing(s) filed on 08 October 2004 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).  
11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).  
a) ☐ All b) ☐ Some \* c) ☐ None of:  
1. ☐ Certified copies of the priority documents have been received.  
2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.  
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- 1) ☒ Notice of References Cited (PTO-892)  
2) ☐ Notice of Draftperson's Patent Drawing Review (PTO-948)  
3) ☒ Information Disclosure Statement(s) (PTO/5508)  
Paper No(s)/Mail Date 12/6/04 10/8/04  
4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date \_\_\_\_\_  
5) ☐ Notice of Informal Patent Application  
6) ☐ Other: \_\_\_\_\_

**DETAILED ACTION**

***Claim Rejections - 35 USC § 102***

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

**Claims 1-7, 11, 12, 14-18 and 20-23 are rejected under 35 U.S.C. 102(b) as being anticipated by Yoon (US 5797888).**

Regarding claim 1, Yoon discloses a tubular body portion [38] defining a lumen [44] therethrough, the tubular body portion having a proximal and distal end [Fig. 1], a distal end portion [22] secured to the distal end of the tubular body portion, the distal end portion including a distal end wall [24, Fig. 2] configured to facilitate passage of a surgical instrument there through [Fig. 7].

Regarding claim 2, Yoon discloses the distal end portion includes an annular side wall [distal end of 22] depending from an outer terminal edge thereof [Fig. 6, edge labeled as 30].

Regarding claim 3, Yoon discloses the distal end portion is made from an elastomeric material [col. 4, ll. 34-41].

Regarding claim 4, Yoon discloses the distal end wall includes an aperture [29].

Regarding claim 5, Yoon discloses the aperture is coaxially aligned with a central longitudinal axis of the tubular body portion [Fig. 2].

Regarding claims 6 and 7, Yoon discloses the distal end portion is secured to the distal end of the tubular body portion such that the annular side wall is capable of at least partially overlapping and completely overlapping the distal end of the tubular body portion [Fig. 6].

Regarding claim 11, Yoon discloses a flange [42] extending from the proximal end of the tubular body portion.

Regarding claim 12, Yoon discloses the distal end wall of the distal end portion is provided with a region of weakened strength [30].

Regarding claim 14, the distal end wall of the distal end portion can be interpreted as being shaped to define a pocket [Fig. 3].

Regarding claim 15, Yoon discloses a hollow elongate cylindrical body [38] including a distal end portion [Fig. 3] terminating in a distal edge [Fig. 3, labeled 40] and a proximal end portion [42], the cylindrical body defining a central longitudinal axis, an elastomeric cap [22] secured to the distal end portion of the cylindrical body [Fig. 3], the cap including a distal end wall [24, Fig. 2] having an outer terminal edge [Fig. 6, edge labeled as 30], the distal end wall including an aperture [29], wherein a center of the aperture is coaxially aligned with the central longitudinal axis [Fig. 6].

Regarding claim 16, Yoon discloses the cylindrical body is configured to receive a surgical instrument there through [Fig. 7].

Regarding claim 17, Yoon discloses a flange [42] extending outward from a proximal terminal edge of the proximal end portion of the cylindrical body.

Regarding claim 18, Yoon discloses the cap [22] is secured to the distal end of the cylindrical body such that the distal end wall of the cap is spaced a distance from the distal terminal edge of the cylindrical body [Fig. 3].

Regarding claim 20, Yoon discloses the distal end wall of the cap is shaped to define a pocket and the aperture is formed in the pocket [Fig. 3].

Regarding claim 21, Yoon discloses providing an instrument introducer [20] including a hollow tubular body [38] having a distal end portion [40] and proximal end portion [42], defining a lumen [44] therebetween, and a resilient cap [22] secured to the distal end of the tubular body, the cap having an aperture [29] formed therein, inserting the distal end of the instrument introducer into a body cavity [co. 6, ll. 20-25], inserting a surgical instrument [62] into the lumen of the tubular body of the instrument introducer through a proximal end of the tubular body [col. 7, ll. 20-24], advancing the surgical instrument through the lumen of the tubular body until a distal end of the surgical instrument projects out through the aperture of the cap, wherein the cap creates a seal around the perimeter the surgical instrument [col. 7, ll. 24-30].

Regarding claim 22, see rejection of claim 21.

Regarding claim 23, Yoon discloses a tubular body portion [38] defining a lumen [44] therethrough, the tubular body portion having a proximal [42] and distal end [40], a distal end portion [22] secured to the distal end of the tubular body portion, the distal end portion including a distal end wall [24], an annular side wall [22] depending from an outer terminal edge thereof [Fig. 6, labeled as 30], wherein the distal end wall of the distal end portion includes an aperture [29] formed therein, and further where the aperture has a smaller diameter than a diameter of the

end wall [size of 29 between Figs. 6 and 7], and wherein the aperture is provided with a region of weakened strength [30].

**Claims 1-20 and 23 are rejected under 35 U.S.C. 102(b) as being anticipated by  
Turkel et al. (US 5792074).**

Regarding claim 1, Turkel discloses a tubular body portion [22] defining a lumen [20] therethrough, the tubular body portion having a proximal and distal end [Fig. 1], a distal end portion [24] secured to the distal end of the tubular body portion, the distal end portion including a distal end wall [24, Fig. 2] configured to facilitate passage of a surgical instrument there through [18 and 12].

Regarding claim 2, Turkel discloses the distal end portion includes an annular side wall [corresponding part of element 24 where 22a is labeled in Fig. 2a] depending from an outer terminal edge thereof [Fig. 2a].

Regarding claim 3, Turkel discloses the distal end portion is made from an elastomeric material [col. 4, ll. 14-15].

Regarding claim 4, Turkel discloses the distal end wall includes an aperture [26].

Regarding claim 5, Turkel discloses the aperture is coaxially aligned with a central longitudinal axis of the tubular body portion [Fig. 2].

Regarding claims 6 and 7, Turkel discloses the distal end portion is secured to the distal end of the tubular body portion such that the annular side wall is capable of at least partially overlapping and completely overlapping the distal end of the tubular body portion [Fig. 2a].

Regarding claim 8, Turkel discloses a proximal terminal edge of the annular side wall of the distal end portion [24] is secured to a distal terminal edge [22a] of the distal end of the tubular body [22] [Fig. 2a].

Regarding claim 9, Turkel discloses the distal end portion is secured to the distal end of the tubular body by gluing [col. 4, ll. 52-54].

Regarding claim 10, Turkel discloses the tubular body portion is fabricated from polypropylene [col. 4, ll. 39-42].

Regarding claim 11, Turkel discloses a flange [32] extending from the proximal end of the tubular body portion.

Regarding claim 12, Turkel discloses the distal end wall of the distal end portion is provided with a region of weakened strength [26].

Regarding claim 13, Turkel discloses the region of weakened strength includes either score lines or reduced thickness [co. 4, ll. 52-64].

Regarding claim 14, the distal end wall of the distal end portion is shaped to define a pocket [Fig. 2a].

Regarding claim 15, Turkel discloses a hollow elongate cylindrical body [22] including a distal end portion [Fig. 2a] terminating in a distal edge [Fig. 2a, labeled 22a] and a proximal end portion [32], the cylindrical body defining a central longitudinal axis, an elastomeric cap [24] secured to the distal end portion of the cylindrical body [Fig. 2a], the cap including a distal end wall [24, Fig. 2a] having an outer terminal edge [Fig. 2a, edge of 24 corresponding to element labeled as 22a], the distal end wall including an aperture [26], wherein a center of the aperture is coaxially aligned with the central longitudinal axis [Fig. 2a].

Regarding claim 16, Turkel discloses the cylindrical body is configured to receive a surgical instrument there through [Fig. 3].

Regarding claim 17, Turkel discloses a flange [32] extending outward from a proximal terminal edge of the proximal end portion of the cylindrical body.

Regarding claim 18, Turkel discloses the cap [24] is secured to the distal end of the cylindrical body such that the distal end wall of the cap is spaced a distance from the distal terminal edge of the cylindrical body [Fig. 2a].

Regarding claim 19, Turkel discloses the cap [24] is secured to the distal end of the cylindrical body such that a proximal terminal edge of the annular side wall is secured to the distal terminal edge [22a] of the cylindrical body [Fig. 2a].

Regarding claim 20, Turkel discloses the distal end wall of the cap is shaped to define a pocket and the aperture is formed in the pocket [Fig. 2a].

Regarding claim 23, Turkel discloses a tubular body portion [22] defining a lumen [20] therethrough, the tubular body portion having a proximal [32] and distal end [22a], a distal end portion [24] secured to the distal end of the tubular body portion, the distal end portion including a distal end wall [24], an annular side wall [corresponding part of element 24 where element 22a is labeled in Fig. 2a] depending from an outer terminal edge thereof [Fig. 2a], wherein the distal end wall of the distal end portion includes an aperture [26] formed therein, and further where the aperture has a smaller diameter than a diameter of the end wall [Fig. 2a], and wherein the aperture is provided with a region of weakened strength [26].



***Conclusion***

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure.

US 4650459 A	USPAT Sheldon; Donald A.	Convolutely wound paper tampon tube
US 20030120224 A1	US-PGPUB Geiser, Kimberly Marie et al.	Feminine care products for the delivery of therapeutic substances
US 6159200 A	USPAT Verdura; Javier et al.	Systems, methods, and instruments for minimally invasive surgery

Any inquiry concerning this communication or earlier communications from the examiner should be directed to VICTORIA W. CHEN whose telephone number is (571)272-3356. The examiner can normally be reached on M-F 8:30-5.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda Dvorak can be reached on (571) 272-4764. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Art Unit: 3739

/Victoria W Chen/

Examiner, Art Unit 3739

/Michael Peffley/

Primary Examiner, Art Unit 3739